



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

Part Number:	PCRS-10F		
Lot.No.:	202401190101		
Specification:	Volume: <input checked="" type="checkbox"/> 0.1ml <input type="checkbox"/> 0.2ml <input type="checkbox"/> 40ul color: <input type="checkbox"/> white <input checked="" type="checkbox"/> clear product: <input type="checkbox"/> plate <input type="checkbox"/> tube <input checked="" type="checkbox"/> strip		
Product Name:	PCR consumable		
Package:	125pcs/box, 10boxes/case		
Manufacturer:	PakGent Bioscience (Suzhou) Co., Ltd.		
Inspection Date:	15th, April, 2024	Report Date	15th, April, 2024
Inspection items and result:			
Cosmetic:	No black spots, no cuts, clear.		
Package:	Accurate quantity and label, inner package without damage.		
Dimension:	OK		
Function test:	Evaporation rate \leq 10%		
Endotoxin:	<0.5EU/ml		
RNase/DNase:	DNase \leq 4×10^{-6} Kunitz units/ μ l; RNase \leq 10^{-9} Kunitz units/ μ l		
PCR Inhibitors:	Not detectable		
Human DNA:	<30pg		
Sterile:	No		
Conclusion: Comply with Specification			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

Title: QA Manager

Signature: Alvin Cao